

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 12505-1 (1988): Tracheostomy tubes, Part:1 Connectors
[MHD 13: Veterinary Hospital Planning and Surgical
Instruments]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



Indian Standard

TRACHEOSTOMY TUBES

PART 1 CONNECTORS

भारतीय मानक

श्वास नली शल्य क्रिया नलिकाएं

भाग 1 संयोजी

UDC 616·231-089·85 : 621·643·41

DESCRIPTORS : MEDICAL EQUIPMENT. ARTIFICIAL BREATHING APPARATUS.
TRACHEOSTOMY TUBES. PIPE FITTINGS. SPECIFICATIONS

© BIS 1990

BUREAU OF INDIAN STANDARDS
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

Indian Standard

TRACHEOSTOMY TUBES

PART 1 CONNECTORS

NATIONAL FOREWORD

This Indian Standard (Part 1) which is identical with ISO 5366/1 : 1986 'Tracheostomy tubes, Part 1 Connectors', issued by the International Organization for Standardization (ISO), was adopted by the Bureau of Indian Standards on 25 April 1988 on the recommendation of the Anaesthetic, Resuscitation and Allied Equipment Sectional Committee (CPDC 13) and approval of the Consumer Products and Medical Instruments Division Council.

The text of the ISO standard has been approved as suitable for publication as Indian Standard without deviations. Certain conventions are, however, not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- a) Wherever the words 'International Standard' appear referring to this standard, they should be read as 'Indian Standard'.
- b) Comma (,) has been used as a decimal marker while in Indian Standards the current practice is to use a point (.) as the decimal marker.

This Indian Standard is in two parts. Other part is:

IS 12505 (Part 2) : 1988 Tracheostomy tubes : Part 2 Basic requirements (*adoption of ISO 5366/2 : 1985*).

CROSS REFERENCE

The technical committee responsible for the preparation of this standard has reviewed the provisions of the following ISO standards and has decided that they are acceptable for use in conjunction with this standard:

ISO 5356/1-1987 Anaesthetic and respiratory equipment — Conical connectors, Part 1 Cones and sockets.

ISO 4135-1979 Anaesthesiology — Vocabulary.

As in the Original Standard, this Page is Intentionally Left Blank

0 Introduction

There are many advantages in harmonizing the requirements for connectors for tracheostomy tubes with those already widely used for tracheal tubes.

The patient connection port of the breathing system of an anaesthetic machine or ventilator is a coaxial 22 mm male/15 mm female conical connector, an important purpose of the smaller connector being to mate with the 15 mm male tracheal tube connector. It is, therefore, advantageous to use the same size of fitting, for tracheostomy tubes used for patients during surgical operations and for those for whom artificial ventilation or other respiratory support may be required.

Inadvertent disconnection of the breathing system from the tracheostomy tube is a well-recognized hazard and it is therefore specified for the larger sizes of tube (6 mm internal diameter and greater) that the male 15 mm conical fitting shall be permanently attached to the tracheostomy tube.

It is also recommended that the design should permit the fitting of a device to make a more secure attachment between the conical fittings if so desired. Such a device may, however, introduce other hazards such as that of accidental extubation. It should, therefore, be as light and compact as possible with any protuberances designed so as to minimize the likelihood of their catching on attendants, hands, surgical dressings or other equipment.

This International Standard recognizes that a permanently attached conical fitting as specified for the larger tubes may, because of its bulk, prove impracticable for use with infants and small children and that there is a wide variety

of acceptable fittings used in clinical paediatric practice. For these reasons it is not a mandatory requirement that the 15 mm conical fitting be permanently attached to the smaller tracheostomy tubes up to and including size 5,5 mm internal diameter. Any other type of connector, however, should either terminate in a 15 mm male conical fitting at its machine end or mate with an adaptor having such a fitting at its machine end.

NOTE — ISO 5366/2 specifies basic requirements for tracheostomy tubes.

1 Scope and field of application

This part of **ISO 5366** specifies basic requirements for connectors for tracheostomy tubes for use on patients undergoing surgical operations and/or those for whom artificial ventilation or other respiratory support may be required.

2 Reference

ISO 5356/1, *Breathing attachments for inhalation anaesthetic apparatus, lung ventilators and resuscitators — Part 1 : Conical fittings and adaptors for breathing systems.*¹⁾

3 Definitions

For the purpose of this part of **ISO 5366**, the following definitions apply.

3.1 tracheostomy tube (tracheotomy) : Tube designed for insertion into the trachea through a tracheostomy (tracheotomy).²⁾

1) At present at the stage of draft.

2) Definition taken from **ISO 4135. Anaesthesiology — Vocabulary.**

3.2 machine end :

(1) End of the tracheostomy tube which is intended to project from the neck of the patient.

(2) End of connector or the adaptor intended to mate with the breathing system of an anaesthetic machine or ventilator.

4 Connectors for tracheostomy tube of size 6 mm internal diameter and greater

4.1 Tracheostomy tubes of size 6 mm internal diameter and greater shall have at the machine end a permanently attached male 15 mm conical fitting in accordance with ISO 5356/1 with the exception that the internal diameter of the conical element shall be not less than the internal diameter of the tube.

4.2 Any transition in internal diameter shall be tapered to give an adequate lead-in for passage of a suction catheter.

5 Connectors for tracheostomy tube of size 5,5 mm internal diameter and less

Connectors for tracheostomy tubes of size 5,5 mm internal diameter and less need not be permanently attached to the tracheostomy tube but shall either

a) have at the machine end a 15 mm male conical fitting,

or

b) mate with an adaptor which shall have a 15 mm male conical fitting in accordance with ISO 5356/1 at its machine end.

6 Retaining devices

6.1 A retaining or locking device may be incorporated in the design to provide added security of attachment of the conical fittings.

6.2 Any projections (for example hooks, lugs or studs) should be so designed as to minimize the risk of their catching on surgical dressings or other equipment.

Bureau of Indian Standards

BIS is a statutory institution established under the Bureau of *Indian Standards Act, 1986* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of **BIS**. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), **BIS**.

Revision of Indian Standards

Indian Standards are reviewed periodically and revised, when necessary and amendments, if any, are issued from time to time. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition. Comments on this Indian Standard may be sent to **BIS** giving the following reference:

Doc : No. CPDC 13 (1910)

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002
Telephones : 331 01 31, 331 13 75

Telegrams : Manaksanstha
(Common to all Offices)

Regional Offices:

Telephone

Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg,
NEW DELHI 110002

{ 331 01 31
331 13 75

Eastern : 1/14 C. I. T. Scheme VII M, V. I. P. Road, Maniktola
CALCUTTA 700054

37 66 62

Northern : SCO 445-446, Sector 35-C, CHANDIGARH 160036

2 18 43

Southern : C. I. T. Campus, IV Cross Road, Madras 600113

41 29 16

Western : Manakalaya, E9 MIDC, Marol, Andheri (East)
BOMBAY 400093

6 32 92 95

Branches : AHMADABAD. BANGALORE. BHOPAL. BHUBANESHWAR.
COIMBATORE. FARIDABAD. GHAZIABAD. GUWAHATI.
HYDERABAD. JAIPUR. KANPUR. PATNA. TRIVANDRUM.